

## QUALITY CERTIFICATE Certificate of Analysis (CoA)

**Ref:** 3020012 - Syringe Water with NaCl 0,9%, Luer Lock, 20 ml-Fill 20 ml

**Batch:** 2411100

**Expiry date:** 2027-07

The following data are related to the sodium chloride (NaCl) 0,9% solution:

Tests:	Specifications:	Results per batch:	
		<b>2406144</b>	<b>2409151</b>
pH	4.5 – 7.0	6.2	6.3
Iron	<1 ppm	<0.04 ppm	0.0411 ppm
Heavy metals	<1 ppm	<0.04 ppm	0.12 ppm
Bacterial Endotoxins	< 0.5 EU/mL	<0.5 EU/mL	<0.5 EU/mL
NaCl Concentration	0.86-0.95 g/100mL	0.90 g/100ml	0.94 g/100ml

The following data are related to the prefilled syringes testing:

Tests:	Specifications:	Results per batch:	
		<b>2407138</b>	<b>2410046</b>
Clarity	Transparent	Complies	Complies
Colour	Colourless	Complies	Complies
Extractable volume	≥ 20.00 mL	20.26 mL	20.46 mL

The following data are related to the prefilled syringes final release testing:

Tests:	Specifications:	Result
Bacterial Endotoxins	< 0.5 EU/mL	<0.5 EU/mL

The above mentioned product prefilled syringe NaCl 0,9% has been manufactured by Sterisets Manufacturing S.A. and complies with the specifications for batch release according to IMP 214 and IMP 075.

### Quality Department

*Joana Pinto*

19/12/2024

### Sterisets Manufacturing S.A.

Zona Industrial 1 – Lote 34 a 35, 4560-164 Guilhufe, Penafiel PORTUGAL

Phone: +31 412 236 606

[www.sterisets.com](http://www.sterisets.com)

PART OF STI GROUP

