

QUALITY CERTIFICATE Certificate of Analysis (CoA)

Ref: 3010010 - Syringe Water with NaCl 0,9%, Luer Lock, 10 ml

Batch: 2412097

Expiry date: 2027-10

The following data are related to the sodium chloride (NaCl) 0,9% solution:

Tests:	Specifications:	Results per batch:		
		2410039	2410040-1	2410041-1
pH	4.5 – 7.0	6.3	6.2	6.3
Iron	<1 ppm	<0.04 ppm	<0.04 ppm	<0.2 ppm
Heavy metals	<1 ppm	<0.04 ppm	<0.04 ppm	<0.2 ppm
Bacterial Endotoxins	< 0.5 EU/mL	<0.5 EU/mL	<0.5 EU/mL	<0.5 EU/mL
NaCl Concentration	0.86-0.95 g/100mL	0.91 g/ 100ml	0.93 g/ 100ml	0.90 g/ 100ml

Tests:	Specifications:	Results per batch:		
		2410089-1	2410089-2	2410095-1
pH	4.5 – 7.0	6.7	6.4	6.3
Iron	<1 ppm	<0.2 ppm	<0.2 ppm	<0.2 ppm
Heavy metals	<1 ppm	<0.2 ppm	<0.2 ppm	<0.2 ppm
Bacterial Endotoxins	< 0.5 EU/mL	<0.5 EU/mL	<0.5 EU/mL	<0.5 EU/mL
NaCl Concentration	0.86-0.95 g/100mL	0.93 g/ 100ml	0.94 g/ 100ml	0.93 g/ 100ml

The following data are related to the prefilled syringes testing:

Tests:	Specifications:	Results per batch:	
		2410134	2410210
Clarity	Transparent	Complies	Complies
Colour	Colourless	Complies	Complies
Extractable volume	≥ 10.00 mL	10.17 mL	10.14 mL

The following data are related to the prefilled syringes final release testing:

Tests:	Specifications:	Result
Bacterial Endotoxins	< 0.5 EU/mL	<0.5 EU/mL

Sterisets Manufacturing S.A.

Zona Industrial 1 – Lote 34 a 35, 4560-164 Guilhufe, Penafiel PORTUGAL

Phone: +31 412 236 606

www.sterisets.com

PART OF STI GROUP



The above mentioned product prefilled syringe NaCl 0,9% has been manufactured by Sterisets Manufacturing S.A. and complies with the specifications for batch release according to IMP 214 and IMP 075.

Quality Department

Joana Pinto

31/01/2025

Sterisets Manufacturing S.A.
Zona Industrial 1 – Lote 34 a 35, 4560-164 Guilhufe, Penafiel PORTUGAL
Phone: +31 412 236 606
www.sterisets.com

PART OF STI GROUP

