

QUALITY CERTIFICATE Certificate of Analysis (CoA)

Ref: 3020012 - Syringe Water with NaCl 0,9%, Luer Lock, 20 ml (fill 20ml)

Batch: 2412094

Expiry date: 2027-11

The following data are related to the sodium chloride (NaCl) 0,9% solution:

Tests:	Specifications:	Results per batch:		
		2410171-2	2410172-2	2410175-1
pH	4.5 – 7.0	6.0	6.0	5.9
Iron	<1 ppm	<0.04 ppm	<0.04 ppm	<0.04 ppm
Heavy metals	<1 ppm	<0.04 ppm	<0.04 ppm	<0.04 ppm
Bacterial Endotoxins	< 0.5 EU/mL	<0.5 EU/mL	<0.5 EU/mL	<0.5 EU/mL
NaCl Concentration	0.86-0.95 g/100mL	0.94 g/ 100ml	0.94 g/ 100ml	0.94 g/ 100ml

The following data are related to the prefilled syringes testing:

Tests:	Specifications:	Results per batch:		
		2411051	2411064	2411084
Clarity	Transparent	Complies	Complies	Complies
Colour	Colourless	Complies	Complies	Complies
Extractable volume	≥ 20.00 mL	20.05 mL	20.08 mL	20.37 mL

The following data are related to the prefilled syringes final release testing:

Tests:	Specifications:	Result
Bacterial Endotoxins	< 0.5 EU/mL	<0.5 EU/mL

The above mentioned product prefilled syringe NaCl 0,9% has been manufactured by Sterisets Manufacturing S.A. and complies with the specifications for batch release according to IMP 214 and IMP 075.

Quality Department

Joana Pinto

21/01/2025

Sterisets Manufacturing S.A.

Zona Industrial 1 – Lote 34 a 35, 4560-164 Guilhufe, Penafiel PORTUGAL

Phone: +31 412 236 606

www.sterisets.com