

## QUALITY CERTIFICATE Certificate of Analysis (CoA)

**Ref:** 3990110 (Former: 14363SNS)- Syringe NaCl 0,9%, Luer Lock, 10 ml SNS

**Batch:** 2412015

**Expiry date:** 2027-10

The following data are related to the sodium chloride (NaCl) 0,9% solution:

Tests:	Specifications:	Results per batch:		
		2410003	2410004	2410039
pH	4.5 – 7.0	6.2	6.3	6.3
Iron	<1 ppm	<0.04 ppm	<0.04 ppm	<0.04 ppm
Heavy metals	<1 ppm	<0.04 ppm	<0.04 ppm	<0.04 ppm
Bacterial Endotoxins	< 0.5 EU/mL	<0.5 EU/mL	<0.5 EU/mL	<0.5 EU/mL
NaCl Concentration	0.86-0.95 g/100mL	0.94 g/ 100ml	0.91 g/ 100ml	0.91 g/ 100ml

The following data are related to the prefilled syringes testing:

Tests:	Specifications:	Results per batch:	
		2410130	2410131
Clarity	Transparent	Complies	Complies
Colour	Colourless	Complies	Complies
Extractable volume	≥ 10.00 mL	10.07 mL	10.18 mL

The following data are related to the prefilled syringes final release testing:

Tests:	Specifications:	Result
Bacterial Endotoxins	< 0.5 EU/mL	<0.5 EU/mL

The above mentioned product prefilled syringe NaCl 0,9% has been manufactured by Sterisets Manufacturing S.A. and complies with the specifications for batch release according to IMP 214 and IMP 075.

### Quality Department

*Joana Pinto*

21/01/2025

### Sterisets Manufacturing S.A.

Zona Industrial 1 – Lote 34 a 35, 4560-164 Guilhufe, Penafiel PORTUGAL

Phone: +31 412 236 606

[www.sterisets.com](http://www.sterisets.com)