

QUALITY CERTIFICATE Certificate of Analysis (CoA)

Ref: 3990110 (Former: 14363SNS)- Syringe NaCl 0,9%, Luer Lock, 10 ml SNS

Batch: 2412096

Expiry date: 2027-11

The following data are related to the sodium chloride (NaCl) 0,9% solution:

Tests:	Specifications:	Results per batch:		
		2410100-2	2410101-2	2410170-1
pH	4.5 – 7.0	6.1	6.0	6.3
Iron	<1 ppm	<0.04 ppm	<0.04 ppm	<0.04 ppm
Heavy metals	<1 ppm	0.0437 ppm	<0.04 ppm	<0.04 ppm
Bacterial Endotoxins	< 0.5 EU/mL	<0.5 EU/mL	<0.5 EU/mL	<0.5 EU/mL
NaCl Concentration	0.86-0.95 g/100mL	0.93 g/ 100ml	0.93 g/ 100ml	0.90 g/ 100ml

Tests:	Specifications:	Results per batch:	
		2410170-2	2410175-1
pH	4.5 – 7.0	6.0	5.9
Iron	<1 ppm	<0.04 ppm	<0.04 ppm
Heavy metals	<1 ppm	<0.04 ppm	<0.04 ppm
Bacterial Endotoxins	< 0.5 EU/mL	<0.5 EU/mL	<0.5 EU/mL
NaCl Concentration	0.86-0.95 g/100mL	0.91 g/ 100ml	0.94 g/ 100ml

The following data are related to the prefilled syringes testing:

Tests:	Specifications:	Results per batch:		
		2411008	2411024	2411040
Clarity	Transparent	Complies	Complies	Complies
Colour	Colourless	Complies	Complies	Complies
Extractable volume	≥ 10.00 mL	10.08 mL	10.13 mL	10.10 mL

The following data are related to the prefilled syringes final release testing:

Tests:	Specifications:	Result
Bacterial Endotoxins	< 0.5 EU/mL	<0.5 EU/mL

Sterisets Manufacturing S.A.

Zona Industrial 1 – Lote 34 a 35, 4560-164 Guilhufe, Penafiel PORTUGAL

Phone: +31 412 236 606

www.sterisets.com

PART OF STI GROUP



The above mentioned product prefilled syringe NaCl 0,9% has been manufactured by Sterisets Manufacturing S.A. and complies with the specifications for batch release according to IMP 214 and IMP 075.

Quality Department

Joana Pinto

23/01/2025

Sterisets Manufacturing S.A.
Zona Industrial 1 – Lote 34 a 35, 4560-164 Guilhufe, Penafiel PORTUGAL
Phone: +31 412 236 606
www.sterisets.com

PART OF STI GROUP

